

RADIONUCLIDE SAFETY DATA SHEET

RADIONUCLIDE: Ce/La-134

FORMS: Soluble

PHYSICAL CHARACTERISTICS

Half-life: 3.16 days / 6.45 minutes

DECAY EMISSIONS

Gammas / X-rays		Betas / Positrons (+) / Electrons		Alphas	
E (keV)	%	E (keV, max)	%	E (keV)	%
511	127.2	27.4	7.9		
604.7	5.04	2104 (+)	1.56		
33	60.1	2709 (+)	62.0		
38	13.28				

Only the four most likely emissions of each type are included. Emission must be > 10 keV and > 1% frequency.

NIH HAZARD CATEGORY

C – level (low hazard): ≤ 7.5 mCi - No DRS protocol required

B – level (moderate hazard): > 7.5 mCi and ≤ 15 mCi - DRS approval required for ceiling increase

A – level (high hazard): > 15 mCi - DRS protocol required

EXTERNAL RADIATION HAZARDS

Gamma dose rate, point source @ 1 foot, 1 mCi:

4.15 mrem/hr

Beta dose rate to skin, point source @ 1 foot, 1 mCi:

173 mrem/hr

Contamination skin dose, uniform deposit of 1 µCi/cm²

4060 mrem/hr

INTERNAL RADIATION HAZARDS

Annual Limit on Intake:

Ingestion: 500 µCi

Inhalation: 700 µCi

The value(s) above indicate the activity taken into the body that would result in either 5 rem to the whole body (CEDE) or 50 rem to an organ or tissue (CDE).

SHIELDING

Gammas/X-rays:

16 mm Pb will decrease exposure by 90%.

Betas/electrons:

12 mm Lucite will stop all betas. Lead may be needed if Bremsstrahlung is detectable through plastic.

DOSIMETRY AND BIOASSAY REQS

Dosimetry may be required when working with Ce/La-134. Fill out Dosimetry Evaluation Form, and DRS will determine if it is required.

Whole body count may be required after spills, contamination, or other abnormal occurrences.

SPECIAL PRECAUTIONS:

1. Contamination monitoring for Ce/La-134 should be performed with a pancake GM probe. Typical GM probe efficiency is 10%.
2. Ce/La-134 should be handled with remote tools when possible.
3. Ensure Ce/La-134 is properly shielded when in storage and use. Shield waste containers containing Ce/La-134.
4. Always wear lab coat, gloves, and protective eyewear when handling Ce/La-134.
5. Source vials must be secured in a locked container when not in use.
6. Ce/La-134 waste must be segregated from long half-life radionuclides (> 120 days). Multi-hazard waste must be segregated from medium and long half-life radionuclides (> 30 days but ≤ 120 days and > 120 days respectively).
7. Ce-134 is likely in secular equilibrium with La-134. Decay emissions, dose rates, meter efficiencies, and shielding thicknesses given use this assumption.

References:

- Brookhaven National Laboratory, National Nuclear Data Center, Decay Radiation Search, Version 3/25/2022.
- National Institutes of Health Division of Radiation Safety, DRS-214 Form - Activity Control System - Cumulative Inventory Ceilings, Revised 10/4/2021.
- Nuclear Regulatory Commission, 10 Code of Federal Regulation Part 20 Appendix B.
- National Institute of Standards and Technology, ESTAR: Stopping Power and Range Tables for Electrons, July 2017
- Nuclear Regulatory Commission, Varskin+ Version 1.0
- Smith, D., Stabin, M. Exposure Rate Constants and Lead Shielding Values for over 1,100 Radionuclides. Health Physics Journal. March 2012.
- Ehs.stanford.edu/reference Radionuclide Fact Sheets